
Midmark M11 - v1.0.5
Total Cycles: 4413

Name: RITTER2
Sterilizer ID: V2313228
03/07/2025 03:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.0 F 0.0
2:00 164.8 F 2.2
4:00 182.1 F 6.2
6:00 205.2 F 11.6
8:00 220.1 F 6.0
10:00 233.0 F 10.9
12:00 231.5 F 7.9
14:00 242.1 F 11.6
16:00 245.1 F 13.3
18:00 257.7 F 20.2
20:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.6 F 29.3
2:00 272.6 F 29.9
3:00 271.9 F 28.9
4:00 272.8 F 29.7

Min 271.6 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 235.3 F 0.0
10:00 226.1 F 0.0
15:00 220.2 F 0.0
20:00 206.1 F 0.0
25:00 187.0 F 0.0
30:00 184.4 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

03/07/2025 04:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____