
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000406
10:30 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:16
STERILIZING 4:42
136.5 C 233kPa 4:42
136.4 C 233kPa 5:12
136.1 C 230kPa 5:42
136.3 C 232kPa 6:12
136.4 C 231kPa 6:42
136.1 C 230kPa 7:12
136.3 C 232kPa 7:42
136.4 C 233kPa 8:12
136.4 C 233kPa 8:12
VENTING 8:12
AIR DRYING 9:07
CYCLE COMPLETE 9:09

Digital Signature #
DF9052980254F256
