
Midmark M11 - v1.0.5
Total Cycles: 1647

Name: RITTER2
Sterilizer ID: V2313228
10/24/2022 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.1 F 0.0
2:00 166.5 F 3.0
4:00 175.9 F 6.8
6:00 204.9 F 11.9
8:00 219.4 F 6.6
10:00 232.0 F 11.8
12:00 233.6 F 9.0
14:00 239.5 F 10.2
16:00 246.5 F 14.0
18:00 258.6 F 20.6
20:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.7
2:00 271.3 F 29.0
3:00 272.7 F 29.9
4:00 272.1 F 29.3

Min 271.2 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 217.5 F 0.0
10:00 231.4 F 0.0
15:00 231.2 F 0.0
20:00 224.7 F 0.0
25:00 228.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

10/24/2022 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____