
Midmark M11 - v1.0.0
Total Cycles: 2436

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2022 04:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.1 F 0.0
2:00 192.9 F 4.3
4:00 199.6 F 9.0
6:00 214.7 F 8.1
8:00 226.7 F 8.6
10:00 237.3 F 10.9
12:00 241.0 F 11.6
14:00 239.2 F 10.1
16:00 253.2 F 17.0
18:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.2 F 29.0
2:00 271.4 F 28.6
3:00 272.3 F 29.0
4:00 271.3 F 28.4

Min 271.3 F 28.3
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 225.2 F 0.0
10:00 252.3 F 0.0
15:00 247.1 F 0.0
20:00 233.8 F 0.0
25:00 217.8 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

10/14/2022 05:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____