
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003204
13:46 23/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:41
136.4 C 220kPa 3:41
136.2 C 223kPa 4:11
136.4 C 223kPa 4:41
136.1 C 223kPa 5:11
136.1 C 220kPa 5:41
136.2 C 223kPa 6:11
136.4 C 223kPa 6:41
136.1 C 218kPa 7:11
136.1 C 218kPa 7:11
VENTING 7:11
AIR DRYING 8:03
CYCLE COMPLETE 8:08

Digital Signature #
5C9191C4D56B85E9
