

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1073

Name: RITTER2  
Sterilizer ID: V2313228  
05/04/2022 02:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.1 F 0.0  
2:00 147.5 F 2.5  
4:00 168.6 F 6.0  
6:00 191.6 F 10.9  
8:00 210.1 F 4.6  
10:00 224.7 F 9.7  
12:00 232.3 F 8.9  
14:00 240.9 F 12.6  
16:00 241.1 F 11.4  
18:00 254.1 F 18.2  
20:00 265.6 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 29.8  
2:00 271.7 F 29.0  
3:00 272.7 F 30.0  
4:00 272.4 F 29.4

Min 271.5 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.7 F 0.0  
5:00 192.9 F 0.0  
10:00 206.0 F 0.0  
15:00 210.0 F 0.0  
20:00 211.6 F 0.0  
25:00 210.6 F 0.0  
30:00 211.3 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 21:45  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:00:07  
CYCLE COMPLETE

05/05/2022 06:49 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02