
Midmark M11 - v1.0.5
Total Cycles: 2082

Name: RITTER2
Sterilizer ID: V2313228
02/23/2023 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.6 F 0.0
2:00 78.5 F 0.0
4:00 80.4 F 0.0
6:00 105.2 F 0.0
8:00 136.9 F 1.2
10:00 168.3 F 3.1
12:00 190.4 F 5.9
14:00 205.9 F 9.6
16:00 221.9 F 11.3
18:00 229.6 F 8.1
20:00 239.9 F 12.7
22:00 237.3 F 10.2
24:00 236.4 F 8.7
26:00 247.7 F 14.6
28:00 258.7 F 20.8
30:00 269.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.3 F 29.3
2:00 272.2 F 29.8
3:00 271.5 F 29.1
4:00 272.4 F 29.5

Min 271.3 F 29.0
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 218.3 F 0.0
10:00 222.7 F 0.0
15:00 219.6 F 0.0
20:00 215.5 F 0.0
25:00 230.3 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

02/23/2023 09:37 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____