
Midmark M11 - v1.0.0
Total Cycles: 4661

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2024 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.7 F 0.0
2:00 190.6 F 3.3
4:00 211.4 F 7.7
6:00 225.7 F 12.1
8:00 223.3 F 6.1
10:00 235.3 F 10.5
12:00 233.7 F 7.7
14:00 243.7 F 12.5
16:00 241.8 F 11.0
18:00 253.9 F 17.2
20:00 264.8 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.7
2:00 271.2 F 28.2
3:00 272.4 F 28.8
4:00 271.3 F 28.1

Min 271.1 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 231.1 F 0.0
10:00 216.4 F 0.0
15:00 214.4 F 0.0
20:00 218.3 F 0.0
25:00 217.4 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

06/19/2024 04:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____