
Midmark M11 - v1.0.0
Total Cycles: 2422

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2022 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.2 F 0.0
2:00 196.6 F 4.1
4:00 197.2 F 8.8
6:00 220.7 F 7.2
8:00 232.5 F 9.2
10:00 240.2 F 9.4
12:00 243.6 F 12.5
14:00 242.0 F 11.2
16:00 255.0 F 18.0
18:00 266.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.8 F 28.5
2:00 272.2 F 28.9
3:00 272.3 F 29.1
4:00 271.6 F 28.4

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 230.8 F 0.0
10:00 245.1 F 0.0
15:00 230.8 F 0.0
20:00 225.5 F 0.0
25:00 223.1 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

10/12/2022 03:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____