
Midmark M11 - v1.0.5
Total Cycles: 4216

Name: RITTER2
Sterilizer ID: V2313228
12/16/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 69.1 F 0.0
4:00 76.5 F 0.0
6:00 98.8 F 0.0
8:00 135.1 F 0.1
10:00 163.1 F 2.3
12:00 186.6 F 5.3
14:00 206.6 F 9.1
16:00 216.0 F 13.0
18:00 228.3 F 6.9
20:00 237.8 F 11.1
22:00 231.9 F 7.4
24:00 245.1 F 13.2
26:00 242.8 F 11.7
28:00 254.9 F 17.9
30:00 265.5 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 28.9
2:00 271.8 F 29.0
3:00 272.1 F 28.5
4:00 272.4 F 29.2

Min 271.3 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 219.6 F 0.0
10:00 208.1 F 0.0
15:00 197.2 F 0.0
20:00 185.5 F 0.0
25:00 181.4 F 0.0
30:00 179.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

12/16/2024 11:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____