

-----  
Midmark M11 - v1.0.5  
Total Cycles: 101

Name: RITTER2  
Sterilizer ID: V2313228  
06/15/2021 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.7 F 0.0  
2:00 110.1 F 0.8  
4:00 127.0 F 3.1  
6:00 157.5 F 6.4  
8:00 182.9 F 10.6  
10:00 199.0 F 4.7  
12:00 216.1 F 8.0  
14:00 229.1 F 12.9  
16:00 232.6 F 10.2  
18:00 237.0 F 9.0  
20:00 246.9 F 14.9  
22:00 259.6 F 21.4  
24:00 269.9 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.6  
1:00 271.7 F 29.9  
2:00 272.2 F 30.2  
3:00 272.9 F 30.5  
4:00 272.7 F 30.7

Min 271.0 F 29.4  
Max 273.6 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.3 F 0.0  
5:00 213.7 F 0.0  
10:00 191.1 F 0.0  
15:00 184.0 F 0.0  
20:00 179.3 F 0.0  
25:00 176.3 F 0.0  
30:00 175.1 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 25:12  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:03:39  
CYCLE COMPLETE

06/15/2021 11:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_