

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 008039  
15:39 07/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:38  
STERILIZING 3:49  
136.4 C 223kPa 3:49  
136.3 C 223kPa 4:19  
136.2 C 221kPa 4:49  
136.4 C 223kPa 5:19  
136.2 C 224kPa 5:49  
136.4 C 222kPa 6:19  
136.4 C 223kPa 6:49  
136.2 C 219kPa 7:19  
136.2 C 219kPa 7:19  
VENTING 7:19  
AIR DRYING 8:08  
CYCLE COMPLETE 38:08

-----  
Digital Signature #  
BEE4FE963B213DE2  
-----