
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001681
14:48 10/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:34
STERILIZING 3:37
136.3 C 220kPa 3:37
136.3 C 225kPa 4:07
136.4 C 225kPa 4:37
136.3 C 225kPa 5:07
136.3 C 224kPa 5:37
136.4 C 226kPa 6:07
136.3 C 224kPa 6:37
136.2 C 223kPa 7:07
136.2 C 223kPa 7:07
VENTING 7:07
AIR DRYING 7:58
CYCLE COMPLETE 8:07

Digital Signature #
F1E0BA8619CDB55C
