
Midmark M11 - v1.0.5
Total Cycles: 3489

Name: RITTER2
Sterilizer ID: V2313228
03/28/2024 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.2 F 0.0
2:00 105.2 F 0.0
4:00 126.2 F 0.1
6:00 166.8 F 2.5
8:00 182.7 F 6.6
10:00 202.1 F 11.1
12:00 217.8 F 4.9
14:00 228.6 F 9.7
16:00 234.8 F 9.5
18:00 241.9 F 12.5
20:00 241.2 F 11.3
22:00 254.2 F 18.0
24:00 265.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.5
2:00 271.5 F 29.0
3:00 272.8 F 29.6
4:00 271.8 F 28.8

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 223.1 F 0.0
10:00 225.2 F 0.0
15:00 219.8 F 0.0
20:00 218.3 F 0.0
25:00 219.6 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:43
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:03:58
CYCLE COMPLETE

03/28/2024 04:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____