
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007842
13:46 09/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:52
STERILIZING 4:57
136.5 C 224kPa 4:57
136.4 C 225kPa 5:27
136.4 C 223kPa 5:57
136.1 C 222kPa 6:27
136.3 C 223kPa 6:57
136.2 C 222kPa 7:27
136.1 C 221kPa 7:57
136.3 C 221kPa 8:27
136.3 C 221kPa 8:27
VENTING 8:27
AIR DRYING 9:19
CYCLE COMPLETE 10:23

Digital Signature #
54582040F286AD20
