
Midmark M11 - v1.0.0
Total Cycles: 1392

Name: RITTER 1
Sterilizer ID: V2178198
11/24/2021 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.1 F 0.0
2:00 182.3 F 1.6
4:00 188.1 F 5.3
6:00 199.8 F 10.1
8:00 220.8 F 5.3
10:00 228.4 F 9.5
12:00 237.0 F 9.4
14:00 242.0 F 11.9
16:00 239.6 F 10.0
18:00 252.8 F 16.6
20:00 264.1 F 23.4
22:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.1
2:00 271.9 F 28.6
3:00 271.7 F 28.7
4:00 272.8 F 29.4

Min 271.0 F 28.1
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 220.2 F 0.0
10:00 224.7 F 0.0
15:00 224.6 F 0.0
20:00 229.4 F 0.0
25:00 229.9 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

11/24/2021 11:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____