
Midmark M11 - v1.0.5
Total Cycles: 202

Name: RITTER2
Sterilizer ID: V2313228
07/13/2021 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.6 F 0.0
2:00 180.5 F 3.3
4:00 188.7 F 7.9
6:00 208.7 F 13.4
8:00 224.8 F 8.5
10:00 234.9 F 10.8
12:00 241.1 F 12.5
14:00 241.9 F 11.9
16:00 255.2 F 18.8
18:00 266.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.5 F 29.5
2:00 272.4 F 30.0
3:00 272.0 F 29.2
4:00 272.8 F 30.0

Min 271.6 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 218.8 F 0.0
10:00 229.7 F 0.0
15:00 232.1 F 0.0
20:00 225.3 F 0.0
25:00 223.1 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

07/13/2021 02:21 PM
ACCEPTED: X
REJECTED:
Operator: U01