
Midmark M11 - v1.0.5
Total Cycles: 2568

Name: RITTER2
Sterilizer ID: V2313228
06/26/2023 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.4 F 0.0
2:00 190.0 F 2.5
4:00 196.5 F 7.3
6:00 214.7 F 12.7
8:00 226.9 F 8.4
10:00 237.6 F 11.6
12:00 241.5 F 12.4
14:00 241.6 F 11.7
16:00 255.0 F 18.6
18:00 266.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.7
2:00 271.6 F 29.0
3:00 272.8 F 29.9
4:00 272.2 F 29.3

Min 271.5 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 213.2 F 0.0
10:00 226.2 F 0.0
15:00 229.5 F 0.0
20:00 229.8 F 0.0
25:00 231.8 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

06/26/2023 02:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____