
Midmark M11 - v1.0.0
Total Cycles: 3503

Name: RITTER 1
Sterilizer ID: V2178198
07/10/2023 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 75.7 F 0.0
4:00 79.4 F 0.0
6:00 97.6 F 0.0
8:00 135.5 F 0.0
10:00 167.9 F 1.8
12:00 185.5 F 4.2
14:00 198.8 F 7.2
16:00 214.3 F 10.6
18:00 223.1 F 8.0
20:00 228.1 F 8.1
22:00 239.2 F 12.5
24:00 236.2 F 9.3
26:00 239.0 F 9.7
28:00 246.3 F 13.4
30:00 257.8 F 19.5
32:00 268.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.1 F 28.4
2:00 272.5 F 28.8
3:00 272.2 F 28.9
4:00 271.6 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 203.8 F 0.0
10:00 190.9 F 0.0
15:00 196.5 F 0.0
20:00 200.1 F 0.0
25:00 205.6 F 0.0
30:00 208.0 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:17
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:11:16
CYCLE COMPLETE

07/10/2023 09:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____