
Midmark M11 - v1.0.0
Total Cycles: 2566

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2022 11:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 174.3 F 3.7
4:00 188.8 F 7.9
6:00 209.9 F 12.9
8:00 222.7 F 7.0
10:00 230.6 F 12.1
12:00 233.8 F 8.6
14:00 241.0 F 10.5
16:00 245.5 F 13.0
18:00 257.5 F 19.7
20:00 268.8 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 272.2 F 29.4
2:00 272.0 F 28.8
3:00 271.7 F 29.1
4:00 272.3 F 29.2

Min 271.3 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 224.4 F 0.0
10:00 231.7 F 0.0
15:00 229.1 F 0.0
20:00 228.5 F 0.0
25:00 231.1 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 21:00
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

11/16/2022 12:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____