

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2561

Name: RITTER2  
Sterilizer ID: V2313228  
06/23/2023 02:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
2:00 191.3 F 4.1  
4:00 206.9 F 9.4  
6:00 222.6 F 8.1  
8:00 234.1 F 11.3  
10:00 234.8 F 9.2  
12:00 237.9 F 9.5  
14:00 249.0 F 15.3  
16:00 261.6 F 22.3  
18:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.7 F 29.7  
2:00 272.0 F 29.0  
3:00 271.8 F 29.4  
4:00 272.6 F 29.7  
  
Min 271.4 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 212.3 F 0.0  
10:00 220.3 F 0.0  
15:00 221.5 F 0.0  
20:00 220.8 F 0.0  
25:00 219.7 F 0.0  
30:00 217.9 F 0.0  
Drying Complete

FILLING: 2:18  
HEATING: 18:26  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:60  
CYCLE COMPLETE

06/23/2023 03:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_