
Midmark M11 - v1.0.5
Total Cycles: 3461

Name: RITTER2
Sterilizer ID: V2313228
03/20/2024 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.1 F 0.0
2:00 97.4 F 0.0
4:00 102.6 F 0.1
6:00 133.9 F 2.2
8:00 168.6 F 5.2
10:00 184.1 F 8.7
12:00 203.7 F 12.6
14:00 212.7 F 5.9
16:00 225.3 F 10.2
18:00 233.8 F 9.1
20:00 239.9 F 12.6
22:00 241.1 F 11.4
24:00 252.9 F 17.5
26:00 263.6 F 23.9
28:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 30.5
2:00 272.1 F 30.1
3:00 272.3 F 30.0
4:00 273.0 F 30.4

Min 270.7 F 29.1
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 238.9 F 0.0
10:00 229.0 F 0.0
15:00 219.9 F 0.0
20:00 222.6 F 0.0
25:00 206.6 F 0.0
30:00 207.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:14
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:06:52
CYCLE COMPLETE

03/20/2024 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____