
Midmark M11 - v1.0.0
Total Cycles: 2344

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2022 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.0 F 0.0
2:00 123.5 F 1.3
4:00 140.4 F 4.5
6:00 168.9 F 8.7
8:00 200.0 F 9.3
10:00 223.3 F 7.4
12:00 233.3 F 12.3
14:00 232.8 F 8.4
16:00 243.1 F 11.5
18:00 242.7 F 12.4
20:00 255.6 F 19.1
22:00 267.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.9 F 29.9
2:00 272.1 F 29.1
3:00 272.7 F 29.8
4:00 272.8 F 29.6

Min 271.8 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 217.4 F 0.0
10:00 224.2 F 0.0
15:00 228.4 F 0.0
20:00 230.2 F 0.0
25:00 229.3 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:51
CYCLE COMPLETE

09/20/2022 01:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____