
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001101
13:14 13/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:17
STERILIZING 4:17
136.4 C 225kPa 4:17
136.4 C 226kPa 4:47
136.4 C 225kPa 5:17
136.4 C 224kPa 5:47
136.4 C 222kPa 6:17
136.4 C 224kPa 6:47
136.2 C 220kPa 7:17
135.9 C 220kPa 7:47
135.9 C 220kPa 7:47
VENTING 7:47
AIR DRYING 8:37
CYCLE COMPLETE 8:38

Digital Signature #
4A414B3C329A462D
