
Midmark M11 - v1.0.5
Total Cycles: 3439

Name: RITTER2
Sterilizer ID: V2313228
03/07/2024 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.7 F 0.0
2:00 149.9 F 1.0
4:00 151.9 F 3.8
6:00 169.1 F 7.7
8:00 193.4 F 12.0
10:00 213.2 F 5.6
12:00 223.4 F 10.1
14:00 231.3 F 8.8
16:00 240.7 F 12.3
18:00 239.7 F 11.0
20:00 252.4 F 17.2
22:00 263.1 F 23.6
24:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 271.4 F 29.7
2:00 272.7 F 29.8
3:00 272.4 F 29.8
4:00 273.0 F 30.4

Min 271.0 F 29.1
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 223.8 F 0.0
10:00 230.1 F 0.0
15:00 214.8 F 0.0
20:00 209.1 F 0.0
25:00 208.6 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:47
CYCLE COMPLETE

03/07/2024 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____