
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.3uS / 0.8ppm
CYCLE NUMBER 001108
13:37 03/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:29
STERILIZING 3:28
136.0 C 220kPa 3:28
136.0 C 222kPa 3:59
136.4 C 224kPa 4:29
135.9 C 221kPa 4:59
136.4 C 225kPa 5:29
136.2 C 224kPa 5:59
136.2 C 223kPa 6:29
136.4 C 226kPa 6:59
136.4 C 226kPa 6:59
VENTING 6:59
AIR DRYING 7:48
CYCLE COMPLETE 14:05

Digital Signature #
1B80B797E3ECCA93
