
Midmark M11 - v1.0.0
Total Cycles: 3456

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2023 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.4 F 0.0
2:00 168.9 F 1.8
4:00 184.7 F 5.7
6:00 206.5 F 10.8
8:00 218.8 F 5.2
10:00 232.3 F 10.8
12:00 231.1 F 7.5
14:00 243.8 F 12.4
16:00 244.1 F 12.4
18:00 256.8 F 19.1
20:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.0
2:00 271.7 F 28.7
3:00 272.5 F 29.0
4:00 271.5 F 28.4

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 221.2 F 0.0
10:00 234.8 F 0.0
15:00 231.1 F 0.0
20:00 227.0 F 0.0
25:00 224.9 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

06/28/2023 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____