
Midmark M11 - v1.0.0
Total Cycles: 3435

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2023 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.2 F 0.0
2:00 194.6 F 3.1
4:00 206.6 F 8.3
6:00 225.3 F 9.7
8:00 231.0 F 9.3
10:00 236.2 F 9.3
12:00 244.2 F 12.9
14:00 243.3 F 11.8
16:00 256.4 F 18.7
18:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.4 F 28.1
3:00 272.7 F 29.0
4:00 271.9 F 28.4

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 222.7 F 0.0
10:00 229.9 F 0.0
15:00 229.8 F 0.0
20:00 228.0 F 0.0
25:00 226.4 F 0.0
30:00 225.0 F 0.0

Drying Complete

FILLING: 1:47
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

06/23/2023 12:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____