
Midmark M11 - v1.0.0
Total Cycles: 774

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2021 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.2 F 0.0
2:00 114.8 F 0.0
4:00 119.4 F 0.1
6:00 148.2 F 0.9
8:00 183.7 F 3.7
10:00 194.4 F 7.1
12:00 209.7 F 10.7
14:00 231.1 F 6.9
16:00 226.6 F 9.6
18:00 239.4 F 10.3
20:00 239.0 F 11.5
22:00 238.3 F 9.3
24:00 250.2 F 15.2
26:00 260.5 F 21.1
28:00 270.7 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.5 F 28.7
2:00 272.8 F 28.8
3:00 272.5 F 29.0
4:00 271.9 F 29.0

Min 271.2 F 28.4
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 233.2 F 0.0
10:00 223.7 F 0.0
15:00 216.2 F 0.0
20:00 214.4 F 0.0
25:00 234.1 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:22
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:07:35
CYCLE COMPLETE

04/07/2021 10:39 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____