

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3434

Name: RITTER2  
Sterilizer ID: V2313228  
03/06/2024 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.0 F 0.0  
2:00 71.2 F 0.0  
4:00 94.2 F 0.0  
6:00 102.6 F 0.8  
8:00 137.6 F 3.2  
10:00 162.7 F 6.2  
12:00 185.6 F 9.7  
14:00 202.1 F 9.0  
16:00 215.0 F 6.1  
18:00 226.1 F 10.0  
20:00 232.4 F 9.9  
22:00 237.3 F 10.7  
24:00 234.8 F 8.1  
26:00 247.9 F 14.6  
28:00 259.0 F 20.7  
30:00 269.2 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.5 F 29.5  
2:00 271.4 F 29.0  
3:00 272.7 F 29.5  
4:00 271.4 F 28.8  
  
Min 271.3 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 203.0 F 0.0  
10:00 192.8 F 0.0  
15:00 189.7 F 0.0  
20:00 186.6 F 0.0  
25:00 184.1 F 0.0  
30:00 182.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:07  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:09:09  
CYCLE COMPLETE

03/06/2024 11:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_