
Midmark M11 - v1.0.0
Total Cycles: 3029

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2023 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.1 F 0.0
2:00 161.2 F 3.0
4:00 184.2 F 6.8
6:00 204.5 F 11.6
8:00 217.5 F 5.2
10:00 228.8 F 10.5
12:00 231.1 F 7.2
14:00 243.2 F 13.0
16:00 242.4 F 11.6
18:00 255.4 F 18.3
20:00 267.0 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 28.9
2:00 271.8 F 28.8
3:00 272.3 F 28.9
4:00 271.7 F 28.7

Min 271.5 F 28.4
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 222.9 F 0.0
10:00 232.9 F 0.0
15:00 239.2 F 0.0
20:00 229.7 F 0.0
25:00 226.5 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:28
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

03/22/2023 04:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____