

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2212

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2023 08:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 80.3 F 0.0  
2:00 85.2 F 0.0  
4:00 117.8 F 0.1  
6:00 148.4 F 0.3  
8:00 176.7 F 0.8  
10:00 205.7 F 4.4  
12:00 224.1 F 8.2  
14:00 232.9 F 12.1  
16:00 226.4 F 6.6  
18:00 237.3 F 11.0  
20:00 234.1 F 7.5  
22:00 244.3 F 13.4  
24:00 242.9 F 12.1  
26:00 254.8 F 18.3  
28:00 265.8 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.9 F 29.5  
2:00 271.4 F 28.9  
3:00 272.8 F 29.5  
4:00 271.5 F 28.8

Min 271.3 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 209.6 F 0.0  
10:00 240.2 F 0.0  
15:00 230.1 F 0.0  
20:00 227.0 F 0.0  
25:00 221.9 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:45  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:07:44  
CYCLE COMPLETE

04/04/2023 09:10 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_