
Midmark M11 - v1.0.0
Total Cycles: 4303

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2024 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.9 F 0.0
2:00 107.1 F 1.0
4:00 127.3 F 3.3
6:00 157.5 F 6.7
8:00 182.4 F 10.7
10:00 210.2 F 3.6
12:00 222.2 F 6.5
14:00 230.3 F 10.5
16:00 231.7 F 7.7
18:00 240.2 F 11.2
20:00 236.1 F 8.6
22:00 249.6 F 14.8
24:00 260.7 F 21.0
26:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.7 F 28.7
2:00 271.8 F 28.4
3:00 272.7 F 28.6
4:00 271.7 F 28.4

Min 271.5 F 28.0
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 232.0 F 0.0
10:00 227.5 F 0.0
15:00 227.0 F 0.0
20:00 213.5 F 0.0
25:00 209.5 F 0.0
30:00 211.7 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 26:45
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

02/14/2024 05:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____