

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001266  
16:32 21/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:14  
STERILIZING 4:29  
136.4 C 224kPa 4:29  
136.3 C 224kPa 4:59  
136.4 C 224kPa 5:29  
136.2 C 223kPa 5:59  
136.4 C 225kPa 6:29  
136.4 C 224kPa 6:59  
136.4 C 225kPa 7:29  
136.1 C 223kPa 7:59  
136.1 C 223kPa 7:59  
VENTING 7:59  
AIR DRYING 8:50  
CYCLE COMPLETE 8:59  
-----

Digital Signature #  
BA3310B1155B8A04  
-----