
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006636
9:22 17/11/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:36
STERILIZING 3:36
136.0 C 220kPa 3:36
136.4 C 220kPa 4:06
136.0 C 220kPa 4:36
136.3 C 222kPa 5:06
136.4 C 221kPa 5:36
136.2 C 220kPa 6:06
136.4 C 222kPa 6:36
136.4 C 221kPa 7:06
136.4 C 221kPa 7:06
VENTING 7:06
AIR DRYING 7:55
CYCLE COMPLETE 7:58

Digital Signature #
9A29F5B2BFF0C94A
