
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002028
14:13 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:00
STERILIZING 3:01
136.4 C 225kPa 3:01
136.2 C 225kPa 3:31
136.3 C 225kPa 4:01
136.3 C 225kPa 4:31
136.1 C 222kPa 5:01
136.4 C 227kPa 5:31
136.4 C 227kPa 6:01
136.4 C 223kPa 6:31
136.4 C 223kPa 6:31
VENTING 6:31
AIR DRYING 7:23
CYCLE COMPLETE 30:15

Digital Signature #
CEEAF1FB024297FB
