

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 000041  
9:14 21/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:46  
PRESSURIZING 3:47  
STERILIZING 4:51  
136.5 C 226kPa 4:51  
136.0 C 220kPa 5:21  
136.2 C 222kPa 5:51  
136.2 C 223kPa 6:21  
136.3 C 223kPa 6:51  
136.1 C 223kPa 7:21  
136.4 C 225kPa 7:51  
136.1 C 221kPa 8:21  
136.1 C 221kPa 8:21  
VENTING 8:21  
AIR DRYING 9:13  
CYCLE COMPLETE 24:13  
-----

Digital Signature #  
D32F1BB1344D252D  
-----