
Midmark M11 - v1.0.5
Total Cycles: 1114

Name: RITTER2
Sterilizer ID: V2313228
05/25/2022 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.9 F 0.0
2:00 165.8 F 2.7
4:00 166.8 F 6.1
6:00 197.9 F 11.0
8:00 214.5 F 5.1
10:00 228.7 F 10.5
12:00 229.6 F 7.2
14:00 241.6 F 13.1
16:00 241.0 F 11.6
18:00 254.3 F 18.2
20:00 266.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 30.1
2:00 272.6 F 29.6
3:00 271.4 F 29.1
4:00 271.7 F 29.7

Min 271.2 F 29.0
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 208.9 F 0.0
10:00 229.3 F 0.0
15:00 229.3 F 0.0
20:00 230.3 F 0.0
25:00 227.0 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

05/25/2022 11:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____