
Midmark M11 - v1.0.0
Total Cycles: 3473

Name: RITTER 1
Sterilizer ID: V2178198
06/30/2023 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.6 F 0.0
2:00 182.6 F 3.6
4:00 200.6 F 8.5
6:00 216.0 F 10.1
8:00 228.8 F 9.3
10:00 235.8 F 9.2
12:00 242.9 F 12.5
14:00 241.5 F 11.1
16:00 254.6 F 17.7
18:00 265.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.8 F 28.9
2:00 271.0 F 28.2
3:00 272.5 F 29.0
4:00 272.8 F 29.2

Min 270.9 F 28.2
Max 273.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 215.5 F 0.0
10:00 226.5 F 0.0
15:00 223.2 F 0.0
20:00 224.3 F 0.0
25:00 227.5 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:42
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

06/30/2023 01:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____