

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1360

Name: RITTER 1  
Sterilizer ID: V2178198  
11/11/2021 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.5 F 0.0  
2:00 175.0 F 2.8  
4:00 198.2 F 7.6  
6:00 224.6 F 10.1  
8:00 229.9 F 9.0  
10:00 238.9 F 9.9  
12:00 242.2 F 12.1  
14:00 240.7 F 10.6  
16:00 254.7 F 17.7  
18:00 266.7 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.3 F 28.7  
2:00 272.3 F 28.8  
3:00 272.1 F 28.4  
4:00 272.4 F 28.9  
  
Min 271.6 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 207.4 F 0.0  
10:00 211.7 F 0.0  
15:00 215.5 F 0.0  
20:00 216.7 F 0.0  
25:00 216.9 F 0.0  
30:00 217.1 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

11/11/2021 11:20 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02