

-----  
Midmark M11 - v1.0.5  
Total Cycles: 776

Name: RITTER2  
Sterilizer ID: V2313228  
02/09/2022 03:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
2:00 178.0 F 2.6  
4:00 189.3 F 7.1  
6:00 210.9 F 13.0  
8:00 224.3 F 8.5  
10:00 235.8 F 10.9  
12:00 242.0 F 12.7  
14:00 242.4 F 12.1  
16:00 256.0 F 19.2  
18:00 267.9 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.0 F 29.2  
2:00 272.5 F 29.6  
3:00 271.8 F 29.5  
4:00 272.0 F 29.2  
  
Min 271.6 F 29.0  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.8 F 0.0  
5:00 203.8 F 0.0  
10:00 211.6 F 0.0  
15:00 223.3 F 0.0  
20:00 219.7 F 0.0  
25:00 219.3 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 2:30  
HEATING: 19:11  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

02/09/2022 04:27 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_