
Midmark M11 - v1.0.0
Total Cycles: 2284

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2022 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.4 F 0.0
2:00 205.1 F 5.4
4:00 223.1 F 12.3
6:00 232.8 F 9.1
8:00 236.0 F 8.7
10:00 243.5 F 11.5
12:00 247.1 F 13.5
14:00 260.4 F 20.8
16:00 272.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.4 F 28.6
2:00 271.8 F 28.3
3:00 272.3 F 28.5
4:00 271.6 F 28.2

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 234.0 F 0.0
10:00 228.9 F 0.0
15:00 227.9 F 0.0
Dry Cycle Aborted

FILLING: 2:47
HEATING: 16:31
STERILIZING: 4:00
VENTING: 3:06
DRYING: 18:15
TOTAL CYCLE: 00:44:38

09/02/2022 02:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____