
Midmark M11 - v1.0.0
Total Cycles: 2101

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2022 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.7 F 0.0
2:00 144.6 F 0.7
4:00 175.5 F 2.7
6:00 187.3 F 6.9
8:00 206.8 F 11.5
10:00 219.7 F 5.1
12:00 232.1 F 9.7
14:00 237.9 F 9.9
16:00 241.1 F 11.6
18:00 238.0 F 9.7
20:00 251.7 F 16.2
22:00 263.1 F 22.8
24:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.2
2:00 271.8 F 28.4
3:00 272.0 F 29.0
4:00 272.4 F 28.9

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 225.2 F 0.0
10:00 226.5 F 0.0
15:00 231.5 F 0.0
20:00 227.6 F 0.0
25:00 232.0 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 24:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

07/07/2022 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____