
Midmark M11 - v1.0.0
Total Cycles: 646

Name: MIDMARK
Sterilizer ID: V2178198
02/01/2021 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.2 F 0.0
2:00 133.1 F 0.0
4:00 178.3 F 0.2
6:00 165.5 F 3.6
8:00 190.5 F 7.7
10:00 210.1 F 12.1
12:00 223.9 F 6.5
14:00 237.7 F 11.6
16:00 236.2 F 8.5
18:00 240.5 F 10.3
20:00 246.6 F 13.2
22:00 259.2 F 20.0
24:00 270.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 271.8 F 28.7
2:00 272.4 F 28.9
3:00 272.5 F 28.8
4:00 272.4 F 28.8

Min 271.5 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 264.5 F 0.0
10:00 241.1 F 0.0
15:00 230.4 F 0.0
Dry Cycle Aborted

FILLING: 0:53
HEATING: 25:22
STERILIZING: 4:00
VENTING: 2:56
DRYING: 18:09
TOTAL CYCLE: 00:51:20

02/01/2021 02:19 PM
ACCEPTED: X
REJECTED:
Operator: U01