
Midmark M11 - v1.0.0
Total Cycles: 2640

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2022 04:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.2 F 0.0
2:00 168.1 F 3.2
4:00 180.5 F 6.8
6:00 205.1 F 11.8
8:00 219.1 F 6.5
10:00 232.8 F 11.7
12:00 233.1 F 8.6
14:00 239.7 F 10.3
16:00 246.0 F 13.4
18:00 258.4 F 20.1
20:00 269.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.2 F 29.0
2:00 271.9 F 29.1
3:00 271.9 F 28.8
4:00 272.3 F 29.4

Min 271.4 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 231.5 F 0.0
10:00 231.1 F 0.0
15:00 228.8 F 0.0
20:00 230.9 F 0.0
25:00 224.3 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

12/05/2022 05:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____