

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1521

Name: RITTER 1  
Sterilizer ID: V2178198  
01/13/2022 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.6 F 0.0  
2:00 159.2 F 1.8  
4:00 169.6 F 5.1  
6:00 186.3 F 9.4  
8:00 209.9 F 7.7  
10:00 219.8 F 7.0  
12:00 232.7 F 11.8  
14:00 232.7 F 8.2  
16:00 242.7 F 11.8  
18:00 244.1 F 12.4  
20:00 256.6 F 18.9  
22:00 267.8 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.7 F 29.2  
2:00 271.7 F 28.4  
3:00 271.7 F 28.7  
4:00 272.7 F 29.2

Min 271.3 F 28.2  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 218.7 F 0.0  
10:00 226.6 F 0.0  
15:00 230.0 F 0.0  
20:00 228.4 F 0.0  
25:00 228.7 F 0.0  
30:00 227.1 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 23:22  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:01:39  
CYCLE COMPLETE

01/13/2022 10:32 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02