
Midmark M11 - v1.0.0
Total Cycles: 3097

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2023 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.8 F 0.0
2:00 188.1 F 2.7
4:00 191.2 F 6.7
6:00 212.9 F 11.4
8:00 224.4 F 6.0
10:00 235.7 F 11.3
12:00 234.5 F 8.2
14:00 241.6 F 11.0
16:00 245.6 F 13.1
18:00 258.3 F 19.9
20:00 269.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.1
2:00 271.8 F 28.5
3:00 271.3 F 28.5
4:00 272.4 F 29.1

Min 271.2 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 221.3 F 0.0
10:00 224.4 F 0.0
15:00 240.7 F 0.0
20:00 226.2 F 0.0
25:00 220.1 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 20:60
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

04/12/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____