
Midmark M11 - v1.0.5
Total Cycles: 3345

Name: RITTER2
Sterilizer ID: V2313228
02/01/2024 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.4 F 0.0
2:00 161.1 F 2.5
4:00 179.9 F 6.2
6:00 200.9 F 11.0
8:00 215.0 F 4.8
10:00 230.5 F 9.9
12:00 233.6 F 8.6
14:00 243.9 F 13.4
16:00 244.0 F 12.8
18:00 257.3 F 19.8
20:00 269.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.6
2:00 271.8 F 28.8
3:00 272.2 F 29.6
4:00 272.3 F 29.4

Min 271.3 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 201.7 F 0.0
10:00 202.5 F 0.0
15:00 200.7 F 0.0
20:00 198.9 F 0.0
25:00 198.5 F 0.0
30:00 196.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

02/01/2024 01:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____