

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 008521  
13:56 04/04/2025  
AIR DRYING ONLY  
60 min. MAXIMUM  
CYCLE START 0:00  
AIR DRYING 0:18  
CYCLE COMPLETE 23:45  
-----

Digital Signature #  
60542DFB4A6C8C4E  
-----