
Midmark M11 - v1.0.5
Total Cycles: 1267

Name: RITTER2
Sterilizer ID: V2313228
07/07/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.4 F 0.0
2:00 179.6 F 3.7
4:00 195.6 F 8.9
6:00 223.7 F 7.9
8:00 232.8 F 9.8
10:00 234.3 F 8.7
12:00 244.1 F 13.5
14:00 243.1 F 12.6
16:00 256.1 F 19.3
18:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.1 F 29.0
2:00 272.5 F 29.3
3:00 272.5 F 29.6
4:00 271.9 F 29.4

Min 271.5 F 29.0
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 230.7 F 0.0
10:00 230.7 F 0.0
15:00 213.7 F 0.0
20:00 214.1 F 0.0
25:00 219.8 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

07/07/2022 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____